

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.52  
Project number 1901274  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant and  
Manufacturer

**CONFIDENTIAL**

Identification of the certified type An **Indicator**  
Type : D39-E

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class  $\textcircled{\text{III}}$  or  $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
7 November 2017



C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.52  
Project number 1901274  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-1901274-01 dated 27 September 2017 that includes 28 pages.

## Characteristics of the indicator:

Configuration	Digital load cells
Accuracy class	Ⓐ or Ⓑ
Weighing range(s)	Single interval
Load cell power supply voltage	5 V DC
Fraction of the maximum permissible error	0
Temperature range	-10 °C / +40 °C
Power supply voltage	110 – 220 V AC 50 Hz
Software identification	Version number: 2