



SCHEDULE

WEIGHT AND MEASURES ACT 1972
WEIGHT AND MEASURES (DUTIES AND POWERS OF THE CUSTODIAN) 1981
(Regulation 11)

Approval No.: ATS-0062-20

CERTIFICATE OF APPROVAL

Description	Non-Automatic Weighing Instruments (Class III)	
Manufacturer Code	MWI 01	
Brand	Indicator	Load Cell
	Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang
Model	MPS-XXXX	
Year of Manufacture	2020	
Technical Specifications	Refer technical specification NMIM2020-1755	

TECHNICAL SPECIFICATIONS

Applicant	: Magoda Weighing Instrument Sdn. Bhd. No. 3 Lorong Mak Mandin1, Mak Mandin Industrial Estate, 13400 Butterworth , Penang	
Issue Date	: 27 October 2020	
Description	: Magoda Platform Scales	
Brand	Indicator	Load Cell
	Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang
Model	: MPS-XXXX	
Class	: III	
Number of Verification	: ≤ 3 000 division	
Scale Interval (n)		
Weighing Range	: Single Interval	
Maximum Capacity	: 30 kg to 10 000 kg	
Scale Interval (e)	: 10 g to 5 kg	

3.0 Pattern Approval Plate (Compulsory)


Manufacturer's mark, or name written in full

Name or mark of applicant

Indication of accuracy class

Pattern approval mark for the instrument

Model number



This plate is verified by NMIM

