



SCHEDULE

**WEIGHT AND MEASURES ACT 1972
WEIGHT AND MEASURES (DUTIES AND POWERS OF THE CUSTODIAN) 1981
(Regulation 11)**

Approval No.: ATS-0062-20

CERTIFICATE OF APPROVAL

Description : Non-Automatic Weighing Instruments (Class III)
Manufacturer Code : MWI 01
Brand :

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Year of Manufacture : 2020
Technical Specifications : Refer technical specification NMIM2020-1755

TECHNICAL SPECIFICATIONS

Applicant : Magoda Weighing Instrument Sdn. Bhd.
No. 3 Lorong Mak Mandin1,
Mak Mandin Industrial Estate,
13400 Butterworth , Penang
Issue Date : 27 October 2020
Description : Magoda Platform Scales
Brand :

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Class : III
Number of Verification Scale Interval (n) : ≤3 000 division
Weighing Range : Single Interval
Maximum Capacity : 30 kg to 10 000 kg
Scale Interval (e) : 10 g to 5 kg

- 3.0 **Pattern Approval Plate (Compulsory)**
 Manufacturer's mark, or name written in full
 Name or mark of applicant
 Indication of accuracy class
 Pattern approval mark for the instrument
 Model number



This plate is verified by NMIM

