



SCHEDULE

WEIGHT AND MEASURES ACT 1972
WEIGHT AND MEASURES (DUTIES AND POWERS OF THE CUSTODIAN) 1981
(Regulation 11)

Approval No.: ATS-0062-20

CERTIFICATE OF APPROVAL

Description : Non-Automatic Weighing Instruments (Class III)
Manufacturer Code : MWI 01
Brand :

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Year of Manufacture : 2020
Technical Specifications : Refer technical specification NMIM2020-1755

TECHNICAL SPECIFICATIONS

Applicant : Magoda Weighing Instrument Sdn. Bhd.
No. 3 Lorong Mak Mandin1,
Mak Mandin Industrial Estate,
13400 Butterworth , Penang
Issue Date : 27 October 2020
Description : Magoda Platform Scales
Brand :

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Class : III
Number of Verification : ≤3 000 division
Scale Interval (n) :
Weighing Range : Single Interval
Maximum Capacity : 30 kg to 10 000 kg
Scale Interval (e) : 10 g to 5 kg

- 3.0 **Pattern Approval Plate (Compulsory)**
Manufacturer's mark, or name written in full
Name or mark of applicant
Indication of accuracy class
Pattern approval mark for the instrument
Model number

Manufacturer	: MWI-01
Applicant	: Magoda Weighing Instrument S/B
Class	: III
Pattern Approval	: ATS-0062-20
Model	: MPS-W300
Max	: 300kg
Min	: 2kg
e	: 100g
S/N.No:	



This plate is verified by NMIM

