



SCHEDULE

**WEIGHT AND MEASURES ACT 1972
WEIGHT AND MEASURES (DUTIES AND POWERS OF THE CUSTODIAN) 1981
(Regulation 11)**

Approval No.: ATS-0062-20

CERTIFICATE OF APPROVAL

Description : Non-Automatic Weighing Instruments (Class III)
Manufacturer Code : MWI 01
Brand :

Indicator	Load Cell
Magoda, Newtron & Herculles	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Year of Manufacture : 2020
Technical Specifications : Refer technical specification NMIM2020-1755


TECHNICAL SPECIFICATIONS

Applicant : Magoda Weighing Instrument Sdn. Bhd.
No. 3 Lorong Mak Mandin1,
Mak Mandin Industrial Estate,
13400 Butterworth , Penang
Issue Date : 27 October 2020
Description : Magoda Platform Scales
Brand :

Indicator	Load Cell
Magoda, Newtron & Herculles	MAVIN, Magoda & Zhonghang

Model : MPS-XXXX
Class : III
Number of Verification : ≤3 000 division
Scale Interval (n) :
Weighing Range : Single Interval
Maximum Capacity : 30 kg to 10 000 kg
Scale Interval (e) : 10 g to 5 kg

3.0 Pattern Approval Plate (Compulsory)
 Manufacturer's mark, or name written in full
 Name or mark of applicant
 Indication of accuracy class
 Pattern approval mark for the instrument
 Model number



This plate is verified by NMIM

