



**SCHEDULE**

**WEIGHT AND MEASURES ACT 1972  
WEIGHT AND MEASURES (DUTIES AND POWERS OF THE CUSTODIAN) 1981  
(Regulation 11)**

Approval No.: ATS-0062-20

**CERTIFICATE OF APPROVAL**

**Description** : Non-Automatic Weighing Instruments (Class III )  
**Manufacturer Code** : MWI 01  
**Brand** : 

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

  
**Model** : MPS-XXXX  
**Year of Manufacture** : 2020  
**Technical Specifications** : Refer technical specification NMIM2020-1755


**TECHNICAL SPECIFICATIONS**

**Applicant** : Magoda Weighing Instrument Sdn. Bhd.  
No. 3 Lorong Mak Mandin1,  
Mak Mandin Industrial Estate,  
13400 Butterworth , Penang  
**Issue Date** : 27 October 2020  
**Description** : Magoda Platform Scales  
**Brand** : 

Indicator	Load Cell
Magoda, Newtron & Hercules	MAVIN, Magoda & Zhonghang

  
**Model** : MPS-XXXX  
**Class** : III  
**Number of Verification** : ≤3 000 division  
**Scale Interval (n)** :  
**Weighing Range** : Single Interval  
**Maximum Capacity** : 30 kg to 10 000 kg  
**Scale Interval (e)** : 10 g to 5 kg

**3.0 Pattern Approval Plate (Compulsory)**  
 Manufacturer's mark, or name written in full  
 Name or mark of applicant  
 Indication of accuracy class  
 Pattern approval mark for the instrument  
 Model number



**This plate is verified by NMIM**

